

NEITZ
Made by Neitz for Your Medical Field



DIRECT OPHTHALMOSCOPE

BX α

Observing Fundus With Minimal Corneal Reflex



Polarized Filter

While observing fundus through the ophthalmoscope, you can insert polarizing filter both in the illumination and observation axes. By this way, corneal reflex is minimized while maintaining observation. The polarizing filter in the observation axis can be rotated so that an optimum balance between brightness of visual field and amount of corneal reflex may be achieved.

Bright Illumination

The combination of an effective condensing lens and mirror reflection system makes the best use of brilliant halogen light source.

Wide Range of Corrective Lenses

Corrective powers range from -36D to +35D in 1D steps, ensuring the sharpest focus of fundus image at all times. The corrective lens wheel is rotatable endlessly, allowing smooth change of powers even when the amount of change is great.

Direct Reading Diopter Indicator

The diopter of the corrective lens is indicated directly even when using high plus or high minus power so that the need to add or subtract the power of the auxiliary lens may be eliminated. The diopter indicator is illuminated from inside, making it possible to read the powers clearly even in the dark room.

Separate Operation for Aperture Dial and Filter Lever

The aperture dial and the filter lever are moved separately so that all apertures can be used with a 4000K filter or a polarizing filter.

Aperture Dial

The aperture dial contains a standard aperture, a small aperture for macula observation, a slit aperture, a concentric scale aperture and a redfree filter aperture.

Filter Lever

With a flip of the filter lever, the 4000K filter and the polarizing filter can be selected.

Aperture Shutter

All optical systems are totally enclosed in the instrument head. When not in use, the aperture shutter prevents dust from infiltrating the head.



S neal Reflex

Polarized Filter

While observing fundus through the ophthalmoscope, you can insert polarizing filter both in the illumination and observation axes. By this way, corneal reflex is minimized while maintaining observation. The polarizing filter in the observation axis can be rotated so that an optimum balance between brightness of visual field and amount of corneal reflex may be achieved.

Bright Illumination

The combination of an effective condensing lens and mirror reflection system makes the best use of brilliant halogen light source.

Wide Range of Corrective Lenses

Corrective powers range from -36D to +35D in 1D steps, ensuring the sharpest focus of fundus image at all times. The corrective lens wheel is rotatable endlessly, allowing smooth change of powers even when the amount of change is great.

Direct Reading Diopter Indicator

The diopter of the corrective lens is indicated directly even when using high plus or high minus power so that the need to add or subtract the power of the auxiliary lens may be eliminated. The diopter indicator is illuminated from inside, making it possible to read the powers clearly even in the dark room.

Separate Operation for Aperture Dial and Filter Lever

The aperture dial and the filter lever are moved separately so that all apertures can be used with a 4000K filter or a polarizing filter.

Aperture Dial

The aperture dial contains a standard aperture, a small aperture for macula observation, a slit aperture, a concentric scale aperture and a redfree filter aperture.

Filter Lever

With a flip of the filter lever, the 4000K filter and the polarizing filter can be selected.

Aperture Shutter

All optical systems are totally enclosed in the instrument head. When not in use, the aperture shutter prevents dust from infiltrating the head.



Halogen Ophthalmoscope BX α -13



- Used with C cell battery handle.
- Contents:
 - Halogen Ophthalmoscope BX α head
 - C cell battery handle
 - Spare bulb
- Bulb: Model No. L-30 (3V, 2W)

Halogen Ophthalmoscope BX α -RC



- Used with rechargeable battery.
- Brightness is twice as great as the battery operated model.
- Rechargeable with Dual Desk Charger RC- II .
- Contents:
 - Halogen Ophthalmoscope BX α head
 - Rechargeable battery handle
 - Spare bulb
- Bulb: Model No. L-29 (4V, 3W)
- Rechargeable Battery: Model No. 1000RS
- Recharger: Dual Desk Charger RC-II (120V or 230V)

Halogen Diagnostic Set BX α -12345FO



- Used for diagnosis of eye, ear, nose and throat.
- Contents:
 - Halogen Ophthalmoscope BX α head
 - Otoscope with 2 specula (10pcs, each of 2.5mm and 40mm, Disposable)
 - Laryngoscope
 - C cell battery handle
 - Spare bulb for Ophthalmoscope
 - Spare bulb for Otoscope, Laryngoscope
- Bulb:
 - Ophthalmoscope: Model No.L-30 (3V, 2W)
 - Otoscope and: Model No.L-69 (3V, 1.7W)
 - Laryngoscope: Model No.L-06 (3V, 0.9W)

Halogen Ophthlmo-Otoscope Set BX α -134FO



- Used for diagnosis of eye, ear and nose.
- Contents:
 - Halogen Ophthalmoscope BX α head
 - Otoscope with 2 specula (10pcs, each of 2.5mm and 4.0mm, Disposable)
 - C cell battery handle
 - Spare bulb for Ophthalmoscope
 - Spare bulb for Otoscope
- Bulb:
 - Ophthalmoscope: Model No.L-30 (3V, 2W)
 - Otoscope: Model No.L-69 (3V, 1.7W)

Fiber Otoscope α -34FO



- Used for diagnosis of ear and nose.
- Contents:
 - Otoscope with 2 specula (10pcs, each of 2.5mm and 4.0mm, Disposable)
 - C cell battery handle
 - Spare bulb
- Bulb: Model No.L-69 (3V, 1.7W)

Distributed in India by

 **SMR OPHTHALMIC PVT. LTD.**

209, Linkway Estate, Link Road, Malad West,
Mumbai - 400 064 INDIA.

☎ (+91) 22 6695 0141 / 2 📠 (+91) 98670 17429

✉ sales@smrindia.com 🌐 Website: www.smrindia.com

NEITZ
Made by Neitz for Your Medical Field

Neitz Instruments Co., Ltd.

4F Ichibancho Court, 15-21, Ichibancho,
Chiyoda-ku, Tokyo 102-0082, JAPAN
Tel (+81) 3-3237-0552
Fax (+81) 3-3237-0554
URL <https://www.neitz.co.jp/en/>

• Specification and appearance are subject to change (for improvement) without notice.